

Supplementary Table 1. Detailed parameters of MRI scanners used in this study

| | Tesla | TR | TE | Thickness | FA | FOV | NEX |
|-----------------------|-------|---------|-----|-----------|------|---------|-----|
| GE Discovery MR750w | 3T | 8.5 | 3.2 | 1.0 | 12.0 | 256×256 | 0 |
| Philips ingenia CX | 3T | 8.2 | 3.7 | 1.0 | 8.0 | 240×240 | 1 |
| Philips ingenia 3T | 3T | 8.1 | 3.7 | 1.0 | 8.0 | 240×240 | 1 |
| GE signa excite | 3T | 6.0 | 1.3 | 1.5 | 20.0 | 220×219 | 0 |
| Siemens skyra | 3T | 1,600.0 | 2.8 | 1.0 | 9.0 | 242×250 | 1 |
| Siemens magnetom trio | 3T | 1,600.0 | 2.8 | 1.0 | 9.0 | 250×250 | 1 |
| Siemens verio | 3T | 1,500.0 | 1.9 | 1.0 | 9.0 | 250×250 | 1 |

MRI, magnetic resonance image; TR, repetition time; TE, echo time; FA, flip angle; FOV, field-of-view; NEX, number of excitations; GE, general electrics.